


Important Safety Information

- Please take the time to familiarise yourself with all the aspects of your HydroTilt and its functions before using the chair.
- Please refer to the CareFlex seating manual for full details of the chair's operation, cleaning and maintenance.
- Do not 'ride' on the push handles or subject them to excessive loads
- Do not sit on the Legrest or subject it to excessive loads
- Do not adjust the seat length whilst a client is sat in the chair
- Always ensure that the client's feet do not drag on the ground when the chair is moved by elevating the Legrest to lift their feet off the floor.
- Always ensure that there are no obstructions under the chair and that there is plenty of space around the chair before using the Tilt-in-Space facility.
- Always apply both brakes when the client is transferred to or from the chair.
- Always try to avoid storing or using the chair in areas where it may obstruct exits or passageways.

YOUR LOCAL CAREFLEX DISTRIBUTOR IS:



CareFlex

HydroTilt Quick Set Up Guide

The HydroTilt is a flexible seating system incorporating patented Water Cell Technology, Reflexion Foam and Tilt-in-Space technology to improve posture and pressure management for individuals with specialist seating requirements.

WARNING

A level of supervision appropriate to the user's abilities should be applied at all times.


Your HydroTilt Chair

When set up correctly:

- The HydroTilt should support the user in a comfortable position.
- The Tilt-in-Space and Leg Elevation features should operate smoothly and precisely.

For your HydroTilt to perform correctly it is essential that you have the correct size of chair in the range to suit the client. See the table below to check sizes to user requirements.


HYDROTILT	Mini	Small	Standard
A. Width between wings	44cm	44cm	50cm
B. Seat width	28/33/38/42	28/33/38/42	40/45/50
C. Overall width	65cm	65cm	73cm
D. Space for Tilt-in-space	25cm	31cm	34cm
E. Upright chair length	93cm	93cm	106cm
F. Upright floor to seat	38/40/43/45	38/40/43/45	52cm
G. Legrest length	32/35/37/39	32/35/37/39	49cm
H. Seat back length	67cm	82cm	80cm
I. Seat Depth	31-43cm	31-43cm	38-53cm



If the client experiences severe discomfort or the chair shows signs of excessive wear, damage or incorrect performance, please discontinue its use and contact CareFlex or your local CareFlex Distributor.

① Every HydroTilt can be identified by its unique batch number.* This can be found behind the top left hand corner of the back cushion (Left hand whilst sitting in the chair) on the identification label stapled to the back board.

**Please quote this number in all correspondence with CareFlex or your CareFlex distributor.*



Looking after your HydroTilt

Care and Maintenance

- When the chair is delivered and every time it is used, check the chair over for signs of damage or excessive wear.
- If your chair shows signs of excessive wear, damage or incorrect performance, please take it out of service immediately and contact CareFlex or your local CareFlex Distributor.
- If the client is experiencing severe discomfort, do not use the chair.
- Take it out of service immediately and contact CareFlex or your local CareFlex Distributor.
- The HydroTilt is upholstered in high quality material which is waterproof and fire retardant. However, do not store or use the chair in damp or wet areas.
- Do not attempt to carry out any structural maintenance or adjustment of your HydroTilt or its accessories without consulting CareFlex or your local CareFlex Distributor.

Full details of chair and accessory set up, along with a guide to care and maintenance for the HydroTilt can be found in the CareFlex Seating Manual, which is available in hard copy on request from CareFlex or is free to download at www.careflex.co.uk.


Look after your HydroTilt and it will provide years of trouble-free service.



⑪ Tray

Your HyroTilt can also be fitted with a tray, which can be used to assist in the client's arm positioning; it should never be used as a restraint.


- **Always ensure that the tray does not interfere with any stoma devices, feeding tubes or similar devices fitted into the client's abdomen.**



⑫ Four Point Harness

The four point harness can be used to assist in the client's positioning; it is not to be used as a restraint.


- The four point harness is held in position with a strap from each corner, two straps attach at the client's waist and two over their shoulders
- These straps are attached to the harness with side-lock buckles, the length of the straps can be adjusted through the buckles.
- **Always ensure that the harness is not over tightened and does not interfere with any medical devices implanted in the client's abdomen.**



⑬ Groin Harness


A groin harness can be used to control the client's position; it is not to be used as a restraint:

- The harness consists of a padded cross shape and webbing straps. Two points of the padded cross shape are attached at the junction between the seat and back cushions.
- When seated in the chair, the two loose ends of the harness should be pulled up over the client's thighs and attached to the straps at the rear of the chair.
- The harness can be adjusted through the webbing straps.



② Brakes


- To locate the brakes, push the chair forward so that the rear castors swing around and the brakes can easily be accessed
- Use your foot to press down on the end of the brake lever to lock the castors
- Gently kick the top of the brake lever to unlock the castors
- **Always brake the castors before transferring the client on or off the chair**



③ Seat Depth

Seat Depth is the most important feature for the correct performance of your chair. When set correctly, the client's buttocks should be positioned right back in the seat and there should be a two finger gap between the client's calves and the front of the seat cushion. To adjust seat depth:


- Locate lever with red handle at the back left-hand side of the chair.
- Twist lever down and pull/push handle on right hand side until seat depth is in desired position.
- Release the lever and pull the handle until you feel the seat "click" positively into position.
- **Never adjust the seat depth when the client is in the chair**



④ Tilt-in-Space

Tilt-in-Space is controlled with the lever on the push handle at the back of the chair:


- Gently pull the lever to release Tilt-in-Space.
- Adjust angle of chair to desired position.
- Release the lever to lock.
- **Do not attempt to raise the client to a standing position when the chair is tilted backwards.**



⑤ Legrest Elevation

The legrest can be elevated for a more comfortable seating position:


- Pull out the lever on the side of the HydroTilt.
- Lift or lower the legrest to the desired angle.
- Release lever to lock the legrest position.
- **Never leave the client unsupervised in the chair with the legrest elevated**



⑥ Legrest Extension and Footrest

The legrest extension and footrest can be attached for clients with a longer lower leg length or who require their feet to be supported; the legrest extension and footrest are fixed into a bracket on the back of the legrest.*


- Pull out the locating pin on the bracket.
- Slide the legrest extension or footrest to adjust it to the desired height.
- Release the locating pin.
- The pin should 'click' into position when the extension or footrest is properly located.



*If your chair does not have a bracket on the legrest one can be easily retro-fitted.

⑦ Headrest


The headrest should always support as much of the client's head as possible. It can be adjusted up and down via the strap which loops around the push handle and is held in place with Velcro.



⑧ Adjusting Seat Width

Seat width can be easily adjusted by swapping the removable arm cushions. To adjust the arms:


- Gently separate the Velcro holding the arms in place and swap the arms for the most appropriate size.
- The semi-circular curve of the cushion must always be fitted to the outside of the chair.
- When adjusting the seat width always allow a small, equal gap either side of the client for movement.



⑨ Lap/Pelvic Strap

The lap/pelvic strap length is adjusted by the two part buckle and a 'tri-glide' adjuster:

- To fasten buckle 'click' two halves together
- To release buckle, press tab in middle of buckle and pull two halves apart
- The Lap/Pelvic Strap is a posture aid; it is not to be used as a restraint. The client must be supervised by suitably trained personnel at all times if they are using a Lap/Pelvic strap.



⑩ Motorised Version

If you own a motorised version of the HydroTilt:

- Tilt-in-Space and Legrest elevation are controlled via the hand controller
- The motorised version is battery operated always ensure that the batteries are properly charged before each use.
- **Please refer to the CareFlex Seating manual for full details of charging your motorised HydroTilt**

